



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/063,315	04/10/2002	Alejandro Berenstein	S63.2-10039	6886

490 7590 02/02/2006

VIDAS, ARRETT & STEINKRAUS, P.A.
6109 BLUE CIRCLE DRIVE
SUITE 2000
MINNETONKA, MN 55343-9185

EXAMINER

WEBB, SARAH K

ART UNIT PAPER NUMBER

3731

DATE MAILED: 02/02/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

e

Office Action Summary	Application No. 10/063,315	Applicant(s) BERENSTEIN ET AL.	
	Examiner Sarah K. Webb	Art Unit 3731	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 29 November 2005.
 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 38-41 and 43-53 is/are pending in the application.
 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
 5) ☐ Claim(s) _____ is/are allowed.
 6) ☒ Claim(s) 38-41 and 43-53 is/are rejected.
 7) ☐ Claim(s) _____ is/are objected to.
 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 a) ☐ All b) ☐ Some * c) ☐ None of:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Objections

1. Claims 38 and 46 are objected to because of the following informalities: the newly added language "the coil segments having curved portions" is unclear. It is unclear whether applicant is simply defining the individual windings of the coil or another structural characteristic of the coil. Appropriate correction is required.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 38-40,44-47,49, and 51-53 are rejected under 35 U.S.C. 102(b) as being anticipated by US Patent No. 5,591,198 to Boyle et al.

Boyle discloses a stent that include a coil section (15) that is longer than two serpentine segments (25,30). The stent can be made of stainless steel (column 4, line 44) and expanded by a balloon (35) (column 4, lines 12-39). Boyle states that the length of the stent can range from 5 mm to 50 mm and the diameter can range from 0.001-0.015 inch (less than 0.4 mm) (column 4, lines 40-42). Boyle is considered to meet the new limitations, since the coil can be described as having "curved portions." Adjacent "portions" are annularly aligned (each winding of the coil has the same position relative to the longitudinal axis) and the portions are longitudinally offset (they are spaced along the longitudinal axis).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

3. Claims 41,43,48, and 50 are rejected under 35 U.S.C. 103(a) as being unpatentable over Boyle in view of US Patent No.6,258,117 to Camrud et al.

Boyle fails to form the stent to be self-expanding. Camrud discloses another multi-section stent and teaches that both balloon expandable and self-expanding stents are well known in the art (column 7, lines 20-30) and gives multiples examples of suitable materials, including steel (column 5, lines 53-60). It would have been obvious to one of ordinary skill in the art at the time the invention was made to form the stent of Boyle to be self-expanding, as taught by Camrud, so that a balloon catheter is not required to deploy the stent.

Response to Arguments

4. Applicant's arguments filed 11/29/05 have been fully considered but they are not persuasive. Applicant's arguments do not comply with 37 CFR 1.111(c) because they do not clearly point out the patentable novelty which he or she thinks the claims present in view of the state of the art disclosed by the references cited or the objections made. Further, they do not show how the amendments avoid such references or objections.

Conclusion

5. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Sarah K. Webb whose telephone number is (571) 272-4706. The examiner can normally be reached on Mon-Fri 8-4:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Anhtuan T. Nguyen can be reached on (571) 272-4963. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3731

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

SKW

1/24/06

Julian W. Woo

JULIAN W. WOO
PRIMARY EXAMINER